

Certificate: alloy-metal-ceramic bond

Optimum compatibility of the products used is a prerequisite for durable, esthetic restorations. Therefore, we approve the metal-ceramic IPS Style® for the processing with the following alloys. The bonding criteria according to ISO 9693-1:2012 apply for the compatibility.

Framework materials: Alloys from Ivoclar Vivadent
Veneering material: IPS Style

The following alloys can be veneered with the metal-ceramic IPS Style:

Alloy	Shade	CTE 25–500 °C
High-gold alloys		
Brite Gold®	rich yellow	14.8
Brite Gold® XH	rich yellow	14.4
Golden Ceramic®	rich yellow	14.6
Aquarius Hard	rich yellow	14.5
Aquarius	rich yellow	14.6
d.SIGN® 98	rich yellow	14.3
BioPorta G	yellow	14.5
Y	yellow	14.6
Aquarius XH	yellow	14.1
Y-2	yellow	15.0
Porta Reflex	yellow	14.3
Porta P6	white	14.0
Porta Geo Ti	yellow	14.1
Sagittarius	white	14.0
Y-Lite	yellow	13.9
Y-1	yellow	14.8
d.SIGN® 96	yellow	14.3
Reduced gold alloys		
Porta Impuls	light yellow	14.0
d.SIGN® 91	white	14.2
Porta SMK 82	white	13.9
W	white	14.2
W-5	white	14.0
Lodestar®	white	14.1
W-3	white	13.9
Leo	white	13.9
W-2	white	14.2
Evolution® Lite	white	14.2
Euro 45	white	14.1

Alloy	Shade	CTE 25–500 °C
Palladium-based alloys		
Simidur S2	white	14.2
Spartan® Plus	white	14.3
Spartan®	white	14.2
Capricorn	white	14.1
d.SIGN® 84	white	13.8
Protocol®	white	13.8
Callisto® 75 Pd	white	13.9
Duo Pal 6	white	14.1
Aries	white	14.7
d.SIGN® 67	white	13.9
d.SIGN® 59	white	14.5
Simidur S1S	white	14.8
Simidur Reflex LC	white	14.8
d.SIGN® 53	white	14.8
W-1	white	15.2
Capricorn 15	white	14.3
Callisto® CPG	white	14.2
Implant alloys		
Porta Implant	white	14.2
Callisto® Implant 78	white	13.9
Callisto® Implant 33	white	14.0
Euro 33 Implant	white	14.3
IS-64	white	14.8
Callisto® Implant 60	white	14.5
Base metal alloys		
Colado® NC	white	14.4
4all®	white	13.9
d.SIGN® 30	white	14.5
Colado® CC	white	14.2
Base metal CAD/CAM alloys		
Colado® CAD CoCr4	white	14.4

The range of available alloys may vary from country to country.

Schaan, September 2017



Dr Thomas Hirt
Chief Technology Officer

Alloy manufacturer:
Ivoclar Vivadent, Inc.
175 Pineview Drive
Amherst, N.Y. 14228, USA
Tel. +1 716 691 0010
Tel. +1 800 533 6825
Fax +1 716 691 2285
www.ivoclarvivadent.us.com

Manufacturer and EU representative:
Ivoclar Vivadent AG
Benderstrasse 2
9494 Schaan
Principality of Liechtenstein
Tel. +423 235 35 35
Fax +423 235 33 60
www.ivoclarvivadent.com

Notified Body:
TÜV Süd Product Service GmbH
Riedlerstrasse 65
D-80339 Munich/Germany