

# Callisto® 75 Pd



## Palladium-based Ceramic Alloy

Callisto® 75 Pd is an economic palladium-based alloy for veneering with ceramic material in the layering and press technique due to its coordinated mechanical and physical properties.

<b>Pd</b> 75.2	<b>Au</b> 2.5	<b>Ag</b> 7.1	<b>Sn</b> 9.3	<b>Ga</b> 4.2	<b>In</b> 1.0	<b>Zn</b> <1.0	<b>Re</b> <1.0	<b>Ru</b> <1.0	<b>Li</b> <1.0
-------------------	------------------	------------------	------------------	------------------	------------------	-------------------	-------------------	-------------------	-------------------

### Advantages

- Wide range of indications from single restorations to long-span bridges
- Excellent physical properties with high strength values, ideally suitable also for the press technique and implant superstructures
- Excellent melting and casting properties for a homogeneous microstructure
- Convenient handling and polishing properties
- Compatible with the most popular veneering ceramics for the layering and the press technique

### Indications

Inlays, onlays, partial crowns, single crowns, telescope and conus crowns, posts, short- and long-span bridges, implant superstructures, partial dentures

### Technical data

Colour	white
Type	4
Density (g/cm <sup>3</sup> )	10.8
Melting range (°C)	1130 – 1296
Casting temperature (°C)	1360 – 1396
CTE 25 - 500 °C	13.9
Oxide firing °C/ min / vacuum	900 / 1 / no vacuum
Vickers hardness	230
0.2 % proof stress (MPa)	500
Modulus of elasticity (MPa)	136.000
Elongation (%)	40.0



Dental restorations by  
H.P. Oss, Austria

# Certificate

**Test material: Callisto® 75 Pd**

Composition in %	Pd	Au	Ag	Ga	Sn	In	Zn	Re	Ru	Li
<b>Callisto® 75 Pd</b>	75.2	2.5	7.1	4.2	9.3	1.0	<1.0	<1.0	<1.0	<1.0

**Manufacturer**

Ivoclar Vivadent Inc., 175 Pineview Drive, Amherst, NY 14228, USA

**Corrosion resistance**

The test was conducted according to the international regulations ISO 22674 and ISO 10271: static immersion test through analytical determination of the metal ion release after a 7-day immersion.

**Result:** The metal ion release after 7 days of immersion was not significant.

**Testing facilities:**

– Texas A&M Health Science Center, Mari Koike, DDS, Ph. D.

**Cytotoxicity**

The Agar Diffusion test determines the biological reactivity of cell culture on test material.

**Result:** The test material is considered non-cytotoxic and meets the requirements of the Agar Diffusion test according to ISO 10993-5.

**Mutagenicity**

An Ames assay was conducted to determine any possible cancer potential.

**Test results:** No mutagenicity potential was found to exist in these alloys.

**Kligman Maximization**

This test evaluated the allergenic potential and/or sensitizing capacity of these alloys.

**Test results:** Based on the standards set by the study protocol, these alloys exhibited no reaction of the challenge (0 % sensitization)

**Sensitivity of oral mucosa**

Test to determine the contact sensitivity of these alloys at the buccal oral mucosa.

**Test results:** No reactions were noted in conjunction with these alloys.

**Testing facility:** Toxikon Corporation, 15 Wiggins Avenue, Bedford, Massachusetts

Amherst, June 2013

Dr. George Tysowsky, D. D. S., M. P. H.  
Vice President-Technology